

Updated Design Notes Of The Regulated Current Arc Modulator For The FNAL H- Source

C.Y. Tan,* D.S. Bollinger, and A.G. Sosa

*Fermi National Accelerator Laboratory,
P.O. Box 500, Batavia, IL 60510-5011, USA.*

(Dated: October 3, 2016)

Abstract

This document updates the feedback circuit shown in the document “Design Notes Of The Regulated Current Arc Modulator For The FNAL H- Source”

* cytan@fnal.gov

I. UPDATE

We have redrawn the printed circuit board (PCB) for the circuit and cleaned up the ratsnest. This has helped fix all the noise problems that we have found in the earlier incarnations of our circuit. See Fig. 1.




FIG. 1. Our “updated” circuit mounted in a case. (a) The power supply filter circuit (b) Two 0.1Ω resistors are wired in series to form R_p . The extra inductances have been removed. (c) The high current IGBT MG300Q1US11 that replaced the MOSFET IXFN120N65X2. (d) The digital/analog part of the circuit that creates the V^+ pulse from the trigger and setting that comes from the control system. (e) The low current feedback part of the circuit.

The updated feedback circuit is shown in Fig. 2. Besides the improved layout of the PCB, other improvements to the circuit are as follows:

1. The MOSFET IXFN120N65X2[1] has been replaced with an IGBT MG300Q1US11[2].

of $0.5 \mu\text{s}$. Note: Although all the IXYS Mosfets and IGBTs that we have used so far claim to have a rise and fall times $< 1 \mu\text{s}$, we have not seen this translate to rise and fall times that are faster than we have observed here.




FIG. 3. (a) shows that the arc voltage still has a decay tail after the arc current is turned off. (b) The delay from the rising edge of the trigger pulse to the falling edge of the arc current is still about $6 \mu\text{s}$ as before. However (c) and (d) shows a dramatic improvement in the fall and rise time of the arc current from 11 to $15 \mu\text{s}$ to ~ 2 to $4 \mu\text{s}$.

The three day performance of the arc modulator is shown in Fig. 4. It is clear from this figure that the arc current has remained steady throughout this period. When we look at the period starting from 0900 hours, 09 Sep 2016 until 0931 hours, 12 Sep 2016, we find that the average and standard deviation of the arc current during this period is $(14.80 \pm 0.06) \text{ A}$. In the mean time, during this period, the arc voltage has increased by $\sim 20 \text{ V}$ and gas pressure has decreased by $\sim 0.5 \times 10^{-6} \text{ Torr}$.




FIG. 4. This is the performance of the arc modulator over three days. The arc current is constant while both the arc voltage and the cube pressure clearly wanders.

Appendix A: The pulse circuit for creating V^+

The subcircuit that we have made for creating V^+ for the feedback loop is shown in Fig. 5. This circuit removes the need for a voltage offset circuit (that we had used in our prior circuits) because the single pole double throw (SPDT) analog switch, DG191[3], has one of its throws connected to ground. This then guarantees that ground is really ground. We set the peak of the V^+ pulse with V_{set} that is at DC on another throw of the SPDT switch.

-
- [1] IXYS. IXFN120N65X2 - 650V, 108A MOSFET. [http://ixapps.ixys.com/DataSheet/DS100690B\(IXFN120N65X2\).pdf](http://ixapps.ixys.com/DataSheet/DS100690B(IXFN120N65X2).pdf), 2016. [Online].
 - [2] Toshiba. MG300Q1US11 - Insulated Gate Bipolar Transistor. <http://pdf.searchdatasheet.com/datasheets-0/toshiba/MG300Q1US11.pdf>. [Online].
 - [3] Harris Semiconductor. DG181 thru DG191. http://pdf.datasheetcatalog.com/datasheet_pdf/intersil/DG182_to_DG191.pdf. [Online].




FIG. 5. The V^+ pulse is guaranteed to be at ground when the trigger pulse is at TTL 0 V because one of the throws of the analog switch, DG191, is connected directly to ground.